

Contribution ID: 41

Type: **not specified**

## A novel technology for particle beam monitoring based on thin silicon sensors

*Tuesday, 6 September 2022 11:25 (25 minutes)*

**Purpose:** An innovative detector based on thin silicon UFSD detectors was developed and characterized by the medical physics group of the University and INFN-Torino for single ion discrimination and ion counting in a therapeutic particle beam.

**Materials and Methods:** Thin silicon sensors were developed, characterized and used in a prototype detector aiming at discriminating and counting single protons or carbon ions of therapeutic beams at a maximum fluence rate of  $108 \text{ p}/(\text{cm}^2 \cdot \text{s})$  with a systematic uncertainty of less than 1%. The sensor is segmented into 146 strips with a sensitive area of  $2.7 \times 2.7 \text{ cm}^2$  to cover the beam cross-section of about 1 cm FWHM at the isocenter.

The detector is read out by six custom ASICs housed on a dedicated frontend board connected to 3 FPGAs (Xilinx Kintex7). A LabVIEW program is used to display online the counting rate of each strip and to store the data for offline analysis. An extensive characterization was performed first in the laboratory with a pulse generator and then with proton and carbon ion beams at the Italian National Center of Oncological Hadron-therapy (CNAO).

**Results:** Data were collected and analyzed at different beam energies at CNAO for carbon ions and protons. The beam profile was studied by measuring the number of measured particles as a function of the strip number and fitting the corresponding distribution with a Gaussian. The beam FWHM measured for different energies is compared with nominal values at the isocenter. Moreover, the counting efficiency was determined by comparing the total number of counts with the delivered number of protons for different beam fluences and energies. Finally, the uniformity over 20 identical spills was studied and was found to be better than 1% independently of the beam energy.

**Conclusions:** The tests performed prove the feasibility of directly measuring the particle rate during treatments. Further studies towards using this technology for beam monitoring in clinical practice require improving the radiation resistance, using finer segmentation and increasing the detector-sensitive area.

**Primary author:** ABUJAMI, Mohammed A. A. (Università degli Studi di Torino - INFN, Torino, Italy)

**Co-authors:** VIGNATI, A. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); DATA, E. (1. National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); MAS MILLAN, F. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); DONETTI, Marco (Centro Nazionale di Adroterapia Oncologica (CNAO), Pavia, Italy); MARTÌ VILLARREAL, O.A. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); CIRIO, R. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); SACCHI, R. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); GARBOLINO, S. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); GIORDANENGO, S. (1. National Institute for Nuclear Physics INFN, Turin division, Turin, Italy); MONACO, V. (National Institute for Nuclear Physics INFN, Turin division, Turin, Italy)

**Presenter:** ABUJAMI, Mohammed A. A. (Università degli Studi di Torino - INFN, Torino, Italy)

**Session Classification:** New experimental developments in hadrontherapy

**Track Classification:** Hadrontherapy and associated technologies